

Voluntary Closing Terminal Device

The voluntary closing terminal device is held open by a rubber band or spring. It closes when the amputee actively closes it with the cable control of the prosthesis.



Photo courtesy of Fillauer Companies



Modern Technology. Traditional Values. Proven Techniques.